



EU DECLARATION OF CONFORMITY

Document no.: TL-DoC-001
Version: 01
Date: 2021-04-16
Place: Hanoi

1/ Manufacturer:

THINH LONG INTERNATIONAL TRADING JOINT STOCK COMPANY

Address: B14 – mixed use commercial center project, public service, office and housing for sale, 108 Nguyen Trai Street., Thuong Dinh Ward., Thanh Xuan District., Hanoi City, Vietnam
www.thinhlongjsc.com

Authorized Representative:

VIENMARK s.r.o.
Address: Za trati 192
Vikýřovice 788 13
Czech Republic

2 /Personal protective equipment:

Product code	Name	Type	Classification
TL 001	Filter half-mask PROMASK	Respirator FFP2	Category III

3/ This declaration of conformity is issued under the sole responsibility of the manufacturer :

THINH LONG INTERNATIONAL TRADING JOINT STOCK COMPANY

4/ Purpose of use:

The personal protective equipment Filter half-mask PROMASK is for protection against particulates. The product is a disposable filtering half mask that provides reliable respiratory protection for the user with at least 95% filtration efficiency against solid and liquid particles without oil. This respirator is intended for use in medical facilities, smog and dusty environments, industrial environments, offices, companies, and other areas of everyday life.

5/ Regulation:

The subject of the declaration described in point 4 is in conformity with the relevant Union harmonization legislation: **Regulation EU No. 2016/425** on personal protective equipment

6/ Harmonized standards used:

References to the relevant harmonized standards used: **ČSN EN ISO 149 + A1: 2009 (EN ISO 149: 2001 + A1: 2009)**

7/ Notified body: 1023

Institut pro testování a certifikaci a.s. Zlín, address: třída Tomáše Bati 299, Louky, 763 02 Zlín, Czech Republic performed an EU-type examination (module B) and issued an EU-type examination certificate.

EU type examination certificate no: 21 0140 T/NB

8/ Personal protective equipment is subject to a type conformity assessment procedure based on internal production control together with supervised product inspections at random intervals (module C2) or type conformity based on quality assurance of the production process (module D) under the supervision of a notified body.

On behalf of the manufacturer



Hoang Thi Cham

Chairman